

TASK 16

# Assess a darkroom for white light leaks

An increasing number of radiographs have a grey fog. Equipment, radiation safety measures and processing have all been checked and found to be satisfactory. Check to see if the problem lies with the darkroom itself.

- a) Select a darkroom.
- b) Carry out the white light leak test.
- c) Answer the following questions in the spaces provided below.

1. Were there any white light leaks? YES/NO
  
2. Briefly describe where you saw light: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_
  
3. Are these white light leaks likely to be a problem? \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_
  
4. What is your recommended action? \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Tutor's comments:** \_\_\_\_\_

Satisfactory/Unsatisfactory

\_\_\_\_\_

Signed \_\_\_\_\_

Date \_\_\_\_\_

Tutor